
Technical Data Sheet		Rosenberger		
RPC-2.92	Mismatch Jack	02K150-055S3		



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface
According to

Mechanically compatible with

IEC 61169-35 RPC-3.50 and SMA

Documents

Application note

AN001 "Calibration Services"

Material and plating

Connector parts Center conductor

Outer conductor Dielectric

Substrate

Material

Plating CuBe

Stainless steel

PS Al_2O_3 Gold, min. 1.27 µm, over nickel

Passivated

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel. : +49 8684 18-0 Email: info@rosenberger.de Page

1/2

Technical Data Sheet		Rosenberger		
RPC-2.92	Mismatch Jack	02K150-055S3		

Electrical data

Frequency range DC to 40 GHz

VSWR 1.1 \pm 0.04, DC to 4 GHz

 1.1 ± 0.05 , 4 GHz to 18 GHz 1.1 ± 0.08 , 18 GHz to 40 GHz

DC Resistance 55 Ω Power handling $\leq 0.5 \text{ W}$

Mechanical data

 $\begin{array}{ll} \text{Mating cycles} & \geq 500 \\ \text{Maximum torque} & 1.70 \text{ Nm} \\ \text{Recommended torque} & 0.90 \text{ Nm} \\ \end{array}$

Gauge 0.00 mm to 0.03 mm

Environmental data

Operating temperature range 1 +20 °C to +26 °C Rated temperature range of use 2 0 °C to +50 °C Storage temperature range -40 °C to +85 °C

RoHS compliant

Declaration of calibration options

Factory Calibration

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, traceable to national / international standards.

Accredited Calibration

Optional this calibration standard can be delivered with an Accredited Calibration (DAkkS) having the highest confidence in the traceability. The DAkkS Calibration Certificate issued reports individual calibration results in a complex format, traceable to national / international standards. Calibration results are reported in a dense data set file. The uncertainties are smaller than in a Factory Calibration.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Packing

Standard 1 pce in box Weight 8.0 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Herbert Babinger 09.12.14 Martin Moder 09.12.14 f00 14-1492 Herbert Babinger 09.12.14	Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
	Herbert Babinger	09.12.14	Martin Moder	09.12.14	f00	14-1492	Herbert Babinger	09.12.14

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel. : +49 8684 18-0 Email : info@rosenberger.de Page 2/2

¹ Temperature range over which these specification are valid.

² This range is underneath and above the operating temperature range, within the mismatch is fully functional and could be used without damage.